


## Volatile Organics (80 ppb Standard) on Rtx®-624 by EPA Method 8260 (extracted ion chromatogram, m/z 59 & m/z 73)

### Peaks

1. *tert*-Butyl alcohol (Q1.59)
2. Methyl *tert*-butyl ether (Q1.73)



Time (min.)

GC\_EV00687

### Column

Rtx®-624, 30 m, 0.25 mm ID, 1.40 µm (cat.# 10968)

### Sample

8260B MegaMix® calibration mix kit (cat.# 30475)  
California oxygenates mix (cat.# 30465)  
VOA calibration mix #1 (ketones) (cat.# 30006)  
8260A Surrogate mix (cat.# 30240)  
8260 Internal standard mix (cat.# 30074)

### Injection

Inj. Temp.: 250 °C

### Purge and Trap

Instrument: O.I. Analytical 4560 with 4551A Autosampler  
Trap Type: #10 (Tenax®/silica gel/carbon molecular sieve)  
Purge: 11 min. @ 20 °C, flow 38 ml/min.

### Desorb Preheat

Temp.: 150 °C

Desorb: 1.0 min. @ 190 °C, flow 32 ml/min.

Bake: 10 min. @ 210 °C

Transfer Line Temp.: 110 °C

### Oven

Oven Temp: 35 °C (hold 7 min.)

Carrier Gas: He, constant flow

Flow Rate: 1.3 mL/min.

Dead Time: 1.47 min. @ 35 °C

Detector: Agilent 5971A GC/MS

Transfer Line Temp.: 280 °C

Tune Type: PFTBA/BFB

Scan Range: 35-260 amu

### Notes

Sample size: 10 mL

Sample temp: 40 °C

Water Management: 110 °C purge, 0 °C desorb, 240 °C bake

6-Port Valve: 110 °C

Sparge Mount: 45 °C

Valve Manifold: 50 °C

### Acknowledgement

Other Conditions: prepurge, preheat, dry purge OFF  
Purge & trap courtesy of O.I. Analytical