


# EU Monitored PCBs and PAHs on Rxi-XLB



Peaks	tr (min)	Conc. (µg/mL)	Column	Standard/Sample
1. Naphthalene	6.134	5	Rxi-XLB, 30 m, 0.25 mm ID, 0.25 µm (cat.# 13723)	2-Methylnaphthalene (cat.# 31285)
2. 2-Methylnaphthalene	7.051	5		SV calibration mix #5/610 PAH (cat.# 31011)
3. Acenaphthylene	8.353	5		PCB congener mix #2 (cat.# 32294)
4. Acenaphthene	8.617	5	Diluent:	Methylene chloride
5. Fluorene	9.435	5	Conc.:	5 µg/mL
6. Phenanthrene	10.946	5	Injection	
7. Anthracene	11.045	5	Inj. Vol.:	1 µL splitless
8. PCB 31	11.412	5	Liner:	Topaz Precision inlet liner, 4.0 mm x 6.3 x 78.5 (cat.# 23305)
9. PCB 28	11.451	5	Inj. Temp.:	240 °C
10. PCB 52	11.799	5	Oven	
11. Fluoranthene	12.868	5	Oven Temp.:	40 °C (hold 1 min) to 120 °C at 30 °C/min to 330 °C at 16 °C/min (hold 4 min)
12. PCB 101	12.902	5	Carrier Gas	He, constant flow
13. Pyrene	13.235	5	Flow Rate:	1 mL/min
14. 2-Methylfluoranthene	13.586	5	Detector	MS
15. PCB 118	13.765	5	Mode:	Scan
16. PCB 153	13.937	5	Scan Program:	
17. PCB 138	14.289	5		
18. PCB 180	15.07	5		
19. Benzo[a]anthracene	15.152	5		
20. Triphenylene	15.197	5	Transfer Line Temp.:	250 °C
21. Chrysene	15.227	5	Source Temp.:	300 °C
22. PCB 194	16.127	5	Quad Temp.:	180 °C
23. Benzo[b]fluoranthene	16.799	5	Instrument	Agilent 7890A GC & 5975C MSD
24. Benzo[k]fluoranthene	16.832	5	Sample Preparation	The sample was in a 2 mL short-cap, screw-thread vial (cat.# 21143) and capped with a short-cap, screw-vial closure (cat.# 24495).
25. Benzo[e]pyrene	17.211	5		
26. Benzo[a]pyrene	17.295	5		
27. Indeno[1,2,3-cd]pyrene	19.161	5		
28. Dibenzo[a,h]anthracene	19.209	5		
29. Benzo[ghi]perylene	19.763	5		

Group	Start Time (min)	Scan Range (amu)	Scan Rate (scans/sec)
1	4	50-550	10