

Hydrocarbon Contamination After Column Maintenance

Column Rxi-5ms, 30 m, 0.32 mm ID, 0.25 μm (cat.# 13424)
Sample No sample injected; instrument ran the method conditions with no injection.
Injection splitless (hold 0.5 min)
Liner: 4.0 mm ID single taper inlet liner w/wool, IP deactivated (cat.# 22406)
Inj. Temp.: 250 $^{\circ}\text{C}$
Purge Flow: 40 mL/min
Oven
Oven Temp.: 50 $^{\circ}\text{C}$ (hold 5 min) to 320 $^{\circ}\text{C}$ at 30 $^{\circ}\text{C}/\text{min}$ (hold 10 min)
Carrier Gas He, constant flow
Flow Rate: 5.0 mL/min
Detector FID @ 330 $^{\circ}\text{C}$

