


QuEChERS Extract of Cannabis on Rxi®-17Sil MS and Rxi®-5ms by GCxGC-TOFMS



Column
 Rxi®-17Sil MS 30 m, 0.25 mm ID, 0.25 µm (cat.# 14123)
 Rxi®-5ms 0.9 m, 0.18 mm ID, 0.18 µm

Injection
 Inj. Vol.: 1 µL split (split ratio 20:1)
 Liner: Premium 4 mm Precision® Inlet Liner w/Wool (cat.# 23305.1)
 Inj. Temp.: 275 °C

Oven
 Oven Temp: Rxi®-17Sil MS: 40 °C (hold 1 min.) to 320 °C at 5 °C/min. (hold 3 min.)
 Rxi®-5ms: 45 °C (hold 1 min.) to 325 °C at 5 °C/min. (hold 3 min.)

Carrier Gas Modulation
 Modulator Temp. Offset: 20 °C
 Second Dimension Separation Time: 2 sec.
 Hot Pulse Time: 0.6 sec.
 Cool Time between Stages: 0.4 sec.

Detector
 Transfer Line Temp.: TOFMS
 310 °C
 Analyzer Type: TOF
 Source Temp.: 250 °C
 Electron Energy: 70 eV
 Mass Defect: -20 mu/100 u
 Solvent Delay Time: 4.5 min.
 Tune Type: PFTBA
 Ionization Mode: EI
 Acquisition Range: 45-500 amu
 Spectral Acquisition Rate: 200 spectra/sec

Instrument Notes
 LECO Pegasus 4D GCxGC-TOFMS

Sample Preparation:
 Sample: Up to 2 g of finely ground cannabis in a 50 mL polypropylene centrifuge tube.

Extraction: Add 10 mL acetonitrile and 10 mL of organic-free water, then shake thoroughly. Soak 1 hour, then aggressively vortex 30 minutes. Add Q-sep™ packet (cat.# 26235), shake 1 minute, then centrifuge 5 minutes at 3,000 g with Q-sep™ 3000 centrifuge (cat.# 26230).