


Rose Oil on Rt-βDEXsa



Peaks	tr 1	tr 2	tr 3	tr 4
1. Rose oxides	22.700	22.920	23.360	23.610
2. (-) Menthone	28.416	—	—	—
3. β-Citronellol (±)	37.066	37.230	—	—
4. Geraniol	39.805	—	—	—

Column Rt-βDEXsa, 30 m, 0.32 mm ID, 0.25 μm (cat.# 13108)
Standard/Sample Rose oil
Diluent: Acetone
Conc.: 5 %
Injection
 Inj. Vol.: 1 μL split (split ratio 100:1)
 Liner: Topaz 4.0 mm ID Precision inlet liner w/ wool (cat.# 23305)
 Inj. Temp.: 210 °C
Oven
 Oven Temp.: 40 °C (hold 1 min) to 230 °C at 2 °C/min (hold 3 min)
Carrier Gas Hz, constant flow
 Linear Velocity: 80 cm/sec @ 40 °C
Detector FID @ 230 °C
 Constant Column +
 Constant Make-up: 51.4 mL/min
 Make-up Gas Type: N₂
 Hydrogen flow: 40 mL/min
 Air flow: 400 mL/min
 Data Rate: 50 Hz
Instrument Agilent 7890A GC
Sample Preparation Rose oil was dissolved in acetone to a final concentration of 5% in 2 mL, screw-thread vials (cat.# 21143), and capped with short-cap, screw-vial closures (cat.# 24495).