


Rose Fragrance on Rt- β DEXsa



Peaks	tr 1	tr 2
1. Menthone	28.416	—
2. β -Citronellol (\pm)	37.066	37.230
3. Geraniol	39.805	—

Column	Rt- β DEXsa, 30 m, 0.32 mm ID, 0.25 μ m (cat.# 13108)
Standard/Sample	Rose fragrance
Diluent:	Acetone
Conc.:	5 %
Injection	
Inj. Vol.:	1 μ L split (split ratio 100:1)
Liner:	Topaz 4.0 mm ID Precision inlet liner w/ wool (cat.# 23305)
Inj. Temp.:	210 °C
Oven	
Oven Temp.:	40 °C (hold 1 min) to 230 °C at 2 °C/min (hold 3 min)
Carrier Gas	H ₂ , constant flow
Linear Velocity:	80 cm/sec @ 40 °C
Detector	FID @ 230 °C
Constant Column +	
Constant Make-up:	51.4 mL/min
Make-up Gas Type:	N ₂
Hydrogen flow:	40 mL/min
Air flow:	400 mL/min
Data Rate:	50 Hz
Instrument	Agilent 7890A GC
Sample Preparation	Rose fragrance was dissolved in acetone to a final concentration of 5% in 2 mL, screw-thread vials (cat.# 21143), and capped with short-cap, screw-vial closures (cat.# 24495).
Notes	Unlike the geranium oil and rose oil (GC_FF1372 and GC_FF1373, respectively), there are no rose oxides present.