


## Riser Oil and Tarball #1 on Rxi-17Sil MS and Rxi-1HT (GCxGC-TOFMS)



GC\_PC1258

<b>Column</b>	Rxi-17Sil MS 60 m, 0.25 mm ID, 0.10 $\mu$ m (cat.# custom) Rxi-1HT 1 m, 0.25 mm ID, 0.10 $\mu$ m (cat.# 13950)	<b>Modulation</b>	Modulator Temp. Offset: 20 $^{\circ}$ C
<b>Sample</b>	Riser pipe oil from Deepwater Horizon oil spill SV internal standard mix (5 ng/ $\mu$ L) (cat.# 31206) Tarball #1	Second Dimension	Separation Time: 2.8 sec Hot Pulse Time: 1.0 sec
Diluent:	Methylene chloride	Cool Time between	Stages: 0.4 sec
Conc.:	10 mg/mL	<b>Detector</b>	MS
<b>Injection</b>	1.0 $\mu$ L split (split ratio 10:1)	Mode:	Transfer Line Temp.: 300 $^{\circ}$ C
Inj. Vol.:	Premium 4 mm Precision liner w/wool (cat.# 23305)	Analyzer Type:	TOF
Liner:	275 $^{\circ}$ C	Source Temp.:	250 $^{\circ}$ C
<b>Oven</b>		Electron Energy:	-70 eV
Oven Temp.:	Rxi-17Sil MS: 40 $^{\circ}$ C (hold 1 min) to 320 $^{\circ}$ C at 2.5 $^{\circ}$ C/min (hold 7 min) Rxi-1HT: 45 $^{\circ}$ C (hold 1 min) to 325 $^{\circ}$ C at 2.5 $^{\circ}$ C/min (hold 7 min)	Mass Defect:	100 mu/100 u
<b>Carrier Gas</b>	He, corrected constant flow (1 mL/min)	Ionization Mode:	EI
		Acquisition Range:	45 to 550 amu
		Spectral Acquisition	Rate: 100 spectra/sec
		<b>Instrument</b>	LECO Pegasus 4D GCxGC-TOFMS